

12/12/2019

Washoe County Water ResourcesCSD
PO Box 11130
Reno, NV 89502
Attn: Ben Jesch

OrderID: 19110766

Dear: Ben Jesch

This is to transmit the attached analytical report. The analytical data and information contained therein was generated using specified or selected methods contained in references, such as Standard Methods for the Examination of Water and Wastewater, online edition, Methods for Determination of Organic Compounds in Drinking Water, EPA-600/4-79-020, and Test Methods for Evaluation of Solid Waste, Physical/Chemical Methods (SW846) Third Edition.

The samples were received by WETLAB-Western Environmental Testing Laboratory in good condition on 11/26/2019. Additional comments are located on page 2 of this report.

If you should have any questions or comments regarding this report, please do not hesitate to call.

Sincerely,



Andy Smith
QA Manager

SPARKS

475 E. Greg Street, Suite 119
Sparks, Nevada 89431
tel (775) 355-0202
fax (775) 355-0817
EPA LAB ID: NV00925 - ELAP No: 2523

ELKO

1084 Lamaille Hwy
Elko, Nevada 89801
tel (775) 777-9933
fax (775) 777-9933
EPA LAB ID: NV00926

LAS VEGAS

3230 Polaris Ave. Suite 4
Las Vegas, Nevada 89102
tel (702) 475-8899
fax (702) 622-2868
EPA LAB ID: NV00932

Western Environmental Testing Laboratory

Report Comments

Washoe County Water Resources CSD - 19110766

Specific Report Comments

None

Report Legend

- B -- Blank contamination; Analyte detected above the method reporting limit in an associated blank
- D -- Due to the sample matrix dilution was required in order to properly detect and report the analyte. The reporting limit has been adjusted accordingly.
- HT -- Sample analyzed beyond the accepted holding time
- J -- The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit. The reported result should be considered an estimate.
- K -- The TPH Diesel Concentration reported here likely includes some heavier TPH Oil hydrocarbons reported in the TPH Diesel range as per EPA 8015.
- L -- The TPH Oil Concentration reported here likely includes some lighter TPH Diesel hydrocarbons reported in the TPH Oil range as per EPA 8015.
- M -- The matrix spike/matrix spike duplicate (MS/MSD) values for the analysis of this parameter were outside acceptance criteria due to probable matrix interference. The reported result should be considered an estimate.
- N -- There was insufficient sample available to perform a spike and/or duplicate on this analytical batch.
- NC -- Not calculated due to matrix interference
- QD -- The sample duplicate or matrix spike duplicate analysis demonstrated sample imprecision. The reported result should be considered an estimate.
- QL -- The result for the laboratory control sample (LCS) was outside WETLAB acceptance criteria and reanalysis was not possible. The reported data should be considered an estimate.
- S -- Surrogate recovery was outside of laboratory acceptance limits due to matrix interference. The associated blank and LCS surrogate recovery was within acceptance limits
- SC -- Spike recovery not calculated. Sample concentration >4X the spike amount; therefore, the spike could not be adequately recovered
- U -- The analyte was analyzed for, but was not detected above the level of the reported sample reporting/quantitation limit. The reported result should be considered an estimate.

General Lab Comments

Per method recommendation (section 4.4), Samples analyzed by methods EPA 300.0 and EPA 300.1 have been filtered prior to analysis.

The following is an interpretation of the results from EPA method 9223B:

A result of zero (0) indicates absence for both coliform and Escherichia coli meaning the water meets the microbiological requirements of the U.S. EPA Safe Drinking Water Act (SDWA). A result of one (1) for either test indicates presence and the water does not meet the SDWA requirements. Waters with positive tests should be disinfected by a certified water treatment operator and retested.

Per federal regulation the holding time for the following parameters in aqueous/water samples is 15 minutes: Residual Chlorine, pH, Dissolved Oxygen, Sulfite.

SPARKS

475 E. Greg Street, Suite 119
Sparks, Nevada 89431
tel (775) 355-0202
fax (775) 355-0817
EPA LAB ID: NV00925 - ELAP No: 2523

ELKO

1084 Lamoille Hwy
Elko, Nevada 89801
tel (775) 777-9933
fax (775) 777-9933
EPA LAB ID: NV00926

LAS VEGAS

3230 Polaris Ave. Suite 4
Las Vegas, Nevada 89102
tel (702) 475-8899
fax (702) 622-2868
EPA LAB ID: NV00932

Western Environmental Testing Laboratory

Analytical Report

Washoe County Water ResourcesCSD

Date Printed: 12/12/2019

PO Box 11130

OrderID: 19110766

Reno, NV 89502

Attn: Ben Jesch

Phone: (775) 954-4612 Fax: NoFax

PO\Project: Swan Lake/io49300

Customer Sample ID: Compton

Collect Date/Time: 11/26/2019 09:45

WETLAB Sample ID: 19110766-001

Receive Date: 11/26/2019 23:15

Analyte	Method	Results	Units	DF	RL	Analyzed	LabID
General Chemistry							
Ammonia, as Nitrogen	SM 4500 NH3 D	0.18	mg/L	1	0.050	12/2/2019	NV00925
Total Phosphorous as P	SM 4500-P E	1.3	mg/L	2	0.040	11/27/2019	NV00925
Total Suspended Solids (TSS)	SM 2540D	18	mg/L	1	10	11/26/2019	NV00925
Biochemical Oxygen Demand	SM 5210B	3.7	mg/L	1	2.0	11/27/2019	NV00925
Total Nitrogen	Calc.	ND	mg/L	1	2.5	12/5/2019	NV00925
Total Dissolved Solids (TDS)	SM 2540C	860	mg/L	1	25	12/2/2019	NV00925
Microbiological Analyses							
Total Coliform (MPN)	SM 9223B (Quantitray)	866.4	MPN/100ml	1	1.0	11/26/2019	NV00925
Escherichia Coli (MPN)	SM 9223B (Quantitray)	2.0	MPN/100ml	1	1.0	11/26/2019	NV00925
Anions by Ion Chromatography							
Chloride	EPA 300.0	150	mg/L	10	10	11/26/2019	NV00925
Nitrate Nitrogen	EPA 300.0	ND D	mg/L	10	1.5	11/26/2019	NV00925
Nitrite Nitrogen	EPA 300.0	ND D	mg/L	10	0.60	11/26/2019	NV00925
Sulfate	EPA 300.0	140	mg/L	10	15	11/26/2019	NV00925
Flow Injection Analyses							
Total Kjeldahl Nitrogen	EPA 351.2	1.8	mg/L	1	0.40	12/5/2019	NV00925
Trace Metals by ICP-OES							
Barium	EPA 200.7	0.062	mg/L	1	0.020	12/11/2019	NV00925
Boron	EPA 200.7	0.29	mg/L	1	0.10	12/11/2019	NV00925
Iron	EPA 200.7	1.5	mg/L	1	0.10	12/11/2019	NV00925
Manganese	EPA 200.7	0.10	mg/L	1	0.010	12/11/2019	NV00925
Nickel	EPA 200.7	ND	mg/L	1	0.030	12/11/2019	NV00925
Trace Metals by ICP-MS							
Arsenic	EPA 200.8	0.027	mg/L	1	0.0050	12/9/2019	NV00925
Sample Preparation							
Trace Metals Digestion	EPA 200.2	Complete		1		12/5/2019	NV00925

DF=Dilution Factor, RL = Reporting Limit (minimum 3X the MDL), ND = Not Detected <RL or <MDL (if listed)

Page 3 of 8

SPARKS

475 E. Greg Street, Suite 119
 Sparks, Nevada 89431
 tel (775) 355-0202
 fax (775) 355-0817
 EPA LAB ID: NV00925 - ELAP No: 2523

ELKO

1084 Lamaille Hwy
 Elko, Nevada 89801
 tel (775) 777-9933
 fax (775) 777-9933
 EPA LAB ID: NV00926

LAS VEGAS

3230 Polaris Ave. Suite 4
 Las Vegas, Nevada 89102
 tel (702) 475-8899
 fax (702) 622-2868
 EPA LAB ID: NV00932

Customer Sample ID: Pompe

Collect Date/Time: 11/26/2019 10:30

WETLAB Sample ID: 19110766-002

Receive Date: 11/26/2019 23:15

Analyte	Method	Results	Units	DF	RL	Analyzed	LabID
General Chemistry							
Ammonia, as Nitrogen	SM 4500 NH3 D	0.10	mg/L	1	0.050	12/2/2019	NV00925
Total Phosphorous as P	SM 4500-P E	0.45	mg/L	1	0.020	11/27/2019	NV00925
Total Suspended Solids (TSS)	SM 2540D	20	mg/L	1	10	11/26/2019	NV00925
Biochemical Oxygen Demand	SM 5210B	2.8	mg/L	1	2.0	11/27/2019	NV00925
Total Nitrogen	Calc.	ND	mg/L	1	2.5	12/5/2019	NV00925
Total Dissolved Solids (TDS)	SM 2540C	770	mg/L	1	25	12/2/2019	NV00925
Microbiological Analyses							
Total Coliform (MPN)	SM 9223B (Quantitray)	118.7	MPN/100ml	1	1.0	11/26/2019	NV00925
Escherichia Coli (MPN)	SM 9223B (Quantitray)	2.0	MPN/100ml	1	1.0	11/26/2019	NV00925
Anions by Ion Chromatography							
Chloride	EPA 300.0	120	mg/L	10	10	11/26/2019	NV00925
Nitrate Nitrogen	EPA 300.0	ND	D mg/L	10	1.5	11/26/2019	NV00925
Nitrite Nitrogen	EPA 300.0	ND	D mg/L	10	0.60	11/26/2019	NV00925
Sulfate	EPA 300.0	100	mg/L	10	15	11/26/2019	NV00925
Flow Injection Analyses							
Total Kjeldahl Nitrogen	EPA 351.2	1.5	mg/L	1	0.40	12/5/2019	NV00925
Trace Metals by ICP-OES							
Barium	EPA 200.7	0.057	mg/L	1	0.020	12/11/2019	NV00925
Boron	EPA 200.7	0.29	mg/L	1	0.10	12/11/2019	NV00925
Iron	EPA 200.7	3.2	mg/L	1	0.10	12/11/2019	NV00925
Manganese	EPA 200.7	0.11	mg/L	1	0.010	12/11/2019	NV00925
Nickel	EPA 200.7	ND	mg/L	1	0.030	12/11/2019	NV00925
Trace Metals by ICP-MS							
Arsenic	EPA 200.8	0.031	mg/L	1	0.0050	12/9/2019	NV00925
Sample Preparation							
Trace Metals Digestion	EPA 200.2	Complete		1		12/5/2019	NV00925

Customer Sample ID: Jean

Collect Date/Time: 11/26/2019 10:10

WETLAB Sample ID: 19110766-003

Receive Date: 11/26/2019 23:15

Analyte	Method	Results	Units	DF	RL	Analyzed	LabID
General Chemistry							
Ammonia, as Nitrogen	SM 4500 NH3 D	0.11	mg/L	1	0.050	12/2/2019	NV00925
Total Phosphorous as P	SM 4500-P E	0.72	mg/L	2	0.040	11/27/2019	NV00925
Total Suspended Solids (TSS)	SM 2540D	15	mg/L	1	10	11/26/2019	NV00925
Biochemical Oxygen Demand	SM 5210B	3.0	mg/L	1	2.0	11/27/2019	NV00925
Total Nitrogen	Calc.	ND	mg/L	1	2.5	12/5/2019	NV00925
Total Dissolved Solids (TDS)	SM 2540C	720	mg/L	1	25	12/2/2019	NV00925
Microbiological Analyses							
Total Coliform (MPN)	SM 9223B (Quantitray)	24.6	MPN/100ml	1	1.0	11/26/2019	NV00925
Escherichia Coli (MPN)	SM 9223B (Quantitray)	2.0	MPN/100ml	1	1.0	11/26/2019	NV00925
Anions by Ion Chromatography							
Chloride	EPA 300.0	140	mg/L	10	10	11/26/2019	NV00925

DF=Dilution Factor, RL = Reporting Limit (minimum 3X the MDL), ND = Not Detected <RL or <MDL (if listed)

Page 4 of 8

SPARKS

475 E. Greg Street, Suite 119
 Sparks, Nevada 89431
 tel (775) 355-0202
 fax (775) 355-0817
 EPA LAB ID: NV00925 - ELAP No: 2523

ELKO

1084 Lamaille Hwy
 Elko, Nevada 89801
 tel (775) 777-9933
 fax (775) 777-9933
 EPA LAB ID: NV00926

LAS VEGAS

3230 Polaris Ave. Suite 4
 Las Vegas, Nevada 89102
 tel (702) 475-8899
 fax (702) 622-2868
 EPA LAB ID: NV00932

Customer Sample ID: Jean

Collect Date/Time: 11/26/2019 10:10

WETLAB Sample ID: 19110766-003

Receive Date: 11/26/2019 23:15

Analyte	Method	Results	Units	DF	RL	Analyzed	LabID
Nitrate Nitrogen	EPA 300.0	ND	D mg/L	10	1.5	11/26/2019	NV00925
Nitrite Nitrogen	EPA 300.0	ND	D mg/L	10	0.60	11/26/2019	NV00925
Sulfate	EPA 300.0	120	mg/L	10	15	11/26/2019	NV00925
Flow Injection Analyses							
Total Kjeldahl Nitrogen	EPA 351.2	1.4	mg/L	1	0.40	12/5/2019	NV00925
Trace Metals by ICP-OES							
Barium	EPA 200.7	0.049	mg/L	1	0.020	12/11/2019	NV00925
Boron	EPA 200.7	0.28	mg/L	1	0.10	12/11/2019	NV00925
Iron	EPA 200.7	2.6	mg/L	1	0.10	12/11/2019	NV00925
Manganese	EPA 200.7	0.070	mg/L	1	0.010	12/11/2019	NV00925
Nickel	EPA 200.7	ND	mg/L	1	0.030	12/11/2019	NV00925
Trace Metals by ICP-MS							
Arsenic	EPA 200.8	0.031	mg/L	1	0.0050	12/9/2019	NV00925
Sample Preparation							
Trace Metals Digestion	EPA 200.2	Complete		1		12/5/2019	NV00925

Customer Sample ID: Jean DUP

Collect Date/Time: 11/26/2019 10:20

WETLAB Sample ID: 19110766-004

Receive Date: 11/26/2019 23:15

Analyte	Method	Results	Units	DF	RL	Analyzed	LabID
General Chemistry							
Ammonia, as Nitrogen	SM 4500 NH3 D	0.11	mg/L	1	0.050	12/2/2019	NV00925
Total Phosphorous as P	SM 4500-P E	0.58	mg/L	1	0.020	11/27/2019	NV00925
Total Suspended Solids (TSS)	SM 2540D	25	mg/L	1	10	11/26/2019	NV00925
Biochemical Oxygen Demand	SM 5210B	2.7	mg/L	1	2.0	11/27/2019	NV00925
Total Nitrogen	Calc.	ND	mg/L	1	2.5	12/5/2019	NV00925
Total Dissolved Solids (TDS)	SM 2540C	750	mg/L	1	25	12/2/2019	NV00925
Microbiological Analyses							
Total Coliform (MPN)	SM 9223B (Quantitray)	124.6	MPN/100ml	1	1.0	11/26/2019	NV00925
Escherichia Coli (MPN)	SM 9223B (Quantitray)	1.0	MPN/100ml	1	1.0	11/26/2019	NV00925
Anions by Ion Chromatography							
Chloride	EPA 300.0	130	mg/L	10	10	11/26/2019	NV00925
Nitrate Nitrogen	EPA 300.0	ND	D mg/L	10	1.5	11/26/2019	NV00925
Nitrite Nitrogen	EPA 300.0	ND	D mg/L	10	0.60	11/26/2019	NV00925
Sulfate	EPA 300.0	110	mg/L	10	15	11/26/2019	NV00925
Flow Injection Analyses							
Total Kjeldahl Nitrogen	EPA 351.2	1.7	mg/L	1	0.40	12/5/2019	NV00925
Trace Metals by ICP-OES							
Barium	EPA 200.7	0.048	mg/L	1	0.020	12/11/2019	NV00925
Boron	EPA 200.7	0.28	mg/L	1	0.10	12/11/2019	NV00925
Iron	EPA 200.7	2.0	mg/L	1	0.10	12/11/2019	NV00925
Manganese	EPA 200.7	0.073	mg/L	1	0.010	12/11/2019	NV00925
Nickel	EPA 200.7	ND	mg/L	1	0.030	12/11/2019	NV00925

DF=Dilution Factor, RL = Reporting Limit (minimum 3X the MDL), ND = Not Detected <RL or <MDL (if listed)

Page 5 of 8

SPARKS

475 E. Greg Street, Suite 119
 Sparks, Nevada 89431
 tel (775) 355-0202
 fax (775) 355-0817
 EPA LAB ID: NV00925 - ELAP No: 2523

ELKO

1084 Lamoille Hwy
 Elko, Nevada 89801
 tel (775) 777-9933
 fax (775) 777-9933
 EPA LAB ID: NV00926

LAS VEGAS

3230 Polaris Ave. Suite 4
 Las Vegas, Nevada 89102
 tel (702) 475-8899
 fax (702) 622-2868
 EPA LAB ID: NV00932

Customer Sample ID: Jean DUP

Collect Date/Time: 11/26/2019 10:20

WETLAB Sample ID: 19110766-004

Receive Date: 11/26/2019 23:15

Analyte	Method	Results	Units	DF	RL	Analyzed	LabID
<u>Trace Metals by ICP-MS</u>							
Arsenic	EPA 200.8	0.030	mg/L	1	0.0050	12/9/2019	NV00925
<u>Sample Preparation</u>							
Trace Metals Digestion	EPA 200.2	Complete		1		12/5/2019	NV00925

SPARKS

475 E. Greg Street, Suite 119
 Sparks, Nevada 89431
 tel (775) 355-0202
 fax (775) 355-0817
 EPA LAB ID: NV00925 - ELAP No: 2523

ELKO

1084 Lamaille Hwy
 Elko, Nevada 89801
 tel (775) 777-9933
 fax (775) 777-9933
 EPA LAB ID: NV00926

LAS VEGAS

3230 Polaris Ave. Suite 4
 Las Vegas, Nevada 89102
 tel (702) 475-8899
 fax (702) 622-2868
 EPA LAB ID: NV00932

Western Environmental Testing Laboratory QC Report

QCBatchID	QCType	Parameter	Method	Result	Actual	% Rec	Units
QC19111147	Blank 1	Chloride	EPA 300.0	ND			mg/L
		Nitrate Nitrogen	EPA 300.0	ND			mg/L
		Nitrite Nitrogen	EPA 300.0	ND			mg/L
		Sulfate	EPA 300.0	ND			mg/L
QC19111160	Blank 1	Total Coliform (MPN)	SM 9223B (Qu	ND			MPN/100ml
		Escherichia Coli (MPN)	SM 9223B (Qu	ND			MPN/100ml
QC19111181	Blank 1	Total Phosphorous as P	SM 4500-P E	ND			mg/L
QC19120021	Blank 1	Biochemical Oxygen Demand	SM 5210B	ND			mg/L
QC19120023	Blank 1	Total Suspended Solids (TSS)	SM 2540D	ND			mg/L
QC19120054	Blank 1	Ammonia, as Nitrogen	SM 4500 NH3	ND			mg/L
QC19120177	Blank 1	Total Dissolved Solids (TDS)	SM 2540C	ND			mg/L
QC19120249	Blank 1	Total Kjeldahl Nitrogen	EPA 351.2	ND			mg/L
QC19120389	Blank 1	Arsenic, Dissolved	EPA 200.8	ND			mg/L
QC19120518	Blank 1	Barium, Dissolved	EPA 200.7	ND			mg/L
		Boron, Dissolved	EPA 200.7	ND			mg/L
		Iron, Dissolved	EPA 200.7	ND			mg/L
		Manganese, Dissolved	EPA 200.7	ND			mg/L
		Nickel, Dissolved	EPA 200.7	ND			mg/L

QCBatchID	QCType	Parameter	Method	Result	Actual	% Rec	Units
QC19111147	LCS 1	Chloride	EPA 300.0	10.4	10.0	104	mg/L
		Nitrate Nitrogen	EPA 300.0	2.05	2.00	103	mg/L
		Nitrite Nitrogen	EPA 300.0	0.513	0.500	103	mg/L
		Sulfate	EPA 300.0	26.2	25.0	105	mg/L
QC19111181	LCS 1	Total Phosphorous as P	SM 4500-P E	0.265	0.250	106	mg/L
QC19120021	LCS 1	Biochemical Oxygen Demand	SM 5210B	181	198	91	mg/L
QC19120023	LCS 1	Total Suspended Solids (TSS)	SM 2540D	197	200	99	mg/L
QC19120023	LCS 2	Total Suspended Solids (TSS)	SM 2540D	196	200	98	mg/L
QC19120054	LCS 1	Ammonia, as Nitrogen	SM 4500 NH3 D	0.940	1.00	94	mg/L
QC19120054	LCS 2	Ammonia, as Nitrogen	SM 4500 NH3 D	0.911	1.00	91	mg/L
QC19120177	LCS 1	Total Dissolved Solids (TDS)	SM 2540C	138	150	92	mg/L
QC19120177	LCS 2	Total Dissolved Solids (TDS)	SM 2540C	149	150	99	mg/L
QC19120249	LCS 1	Total Kjeldahl Nitrogen	EPA 351.2	0.956	1.00	96	mg/L
QC19120389	LCS 1	Arsenic, Dissolved	EPA 200.8	0.0451	0.050	90	mg/L
QC19120518	LCS 1	Barium, Dissolved	EPA 200.7	0.954	1.00	95	mg/L
		Boron, Dissolved	EPA 200.7	0.906	1.00	91	mg/L
		Iron, Dissolved	EPA 200.7	0.956	1.00	96	mg/L
		Manganese, Dissolved	EPA 200.7	0.933	1.00	93	mg/L
		Nickel, Dissolved	EPA 200.7	4.71	5.00	94	mg/L

QCBatchID	QCType	Parameter	Method	Duplicate Sample	Sample Result	Duplicate Result	Units	RPD
QC19120023	Duplicate 1	Total Suspended Solids (TSS)	SM 2540D	19110761-001	ND	ND	mg/L	21 %
QC19120023	Duplicate 2	Total Suspended Solids (TSS)	SM 2540D	19110766-004	25.0	29.3	mg/L	16 %
QC19120177	Duplicate 1	Total Dissolved Solids (TDS)	SM 2540C	19110766-001	858	858	mg/L	<1%
QC19120177	Duplicate 2	Total Dissolved Solids (TDS)	SM 2540C	19110766-003	716	736	mg/L	3 %

DF=Dilution Factor, RL = Reporting Limit (minimum 3X the MDL), ND = Not Detected <RL or <MDL (if listed)

Page 7 of 8

SPARKS

475 E. Greg Street, Suite 119
Sparks, Nevada 89431
tel (775) 355-0202
fax (775) 355-0817
EPA LAB ID: NV00925 - ELAP No: 2523

ELKO

1084 Lamoille Hwy
Elko, Nevada 89801
tel (775) 777-9933
fax (775) 777-9933
EPA LAB ID: NV00926

LAS VEGAS

3230 Polaris Ave. Suite 4
Las Vegas, Nevada 89102
tel (702) 475-8899
fax (702) 622-2868
EPA LAB ID: NV00932

QCBatchID	QCType	Parameter	Method	Spike Sample	Sample Result	MS Result	MSD Result	Spike Value	Units	MS %Rec	MSD %Rec	RPD %
QC19111147	MS 1	Chloride	EPA 300.0	19110751-004	2.54	7.56	7.86	5	mg/L	100	106	4
		Nitrate Nitrogen	EPA 300.0	19110751-004	ND	2.03	2.17	2	mg/L	100	107	7
		Nitrite Nitrogen	EPA 300.0	19110751-004	ND	0.499	0.527	0.5	mg/L	99	104	6
		Sulfate	EPA 300.0	19110751-004	8.62	18.3	19.2	10	mg/L	97	106	5
QC19111147	MS 2	Chloride	EPA 300.0	19110721-001	ND	5.49	5.60	5	mg/L	108	111	2
		Nitrate Nitrogen	EPA 300.0	19110721-001	ND	2.18	2.25	2	mg/L	109	112	3
		Nitrite Nitrogen	EPA 300.0	19110721-001	ND	0.544	0.558	0.5	mg/L	109	112	2
		Sulfate	EPA 300.0	19110721-001	ND	11.0	11.2	10	mg/L	107	109	2
QC19111181	MS 1	Total Phosphorous as P	SM 4500-P E	19110762-001	5.24	7.76	7.88	0.25	mg/L	101	105	2
QC19111181	MS 2	Total Phosphorous as P	SM 4500-P E	19110786-013	0.074	0.309	0.277	0.25	mg/L	94	81	11
QC19120054	MS 1	Ammonia, as Nitrogen	SM 4500 NH3	19110766-001	0.177	1.05	1.04	1	mg/L	88	86	1
QC19120054	MS 2	Ammonia, as Nitrogen	SM 4500 NH3	19110785-007	0.240	1.06	1.07	1	mg/L	82	83	<1
QC19120054	MS 3	Ammonia, as Nitrogen	SM 4500 NH3	19110815-001	0.141	M 0.937	0.948	1	mg/L	NC	NC	NC
QC19120249	MS 1	Total Kjeldahl Nitrogen	EPA 351.2	19110761-001	1.01	QD 1.97	2.66	1	mg/L	96	165	30
QC19120249	MS 2	Total Kjeldahl Nitrogen	EPA 351.2	19110766-002	1.52	2.48	2.46	1	mg/L	96	94	<1
QC19120389	MS 1	Arsenic, Dissolved	EPA 200.8	19110757-010	ND	D 0.0636	0.0551	0.05	mg/L	116	22	14
QC19120518	MS 1	Barium, Dissolved	EPA 200.7	19110757-010	ND	D 0.782	0.831	1	mg/L	77	82	6
		Boron, Dissolved	EPA 200.7	19110757-010	ND	D 1.03	0.953	1	mg/L	103	95	8
		Iron, Dissolved	EPA 200.7	19110757-010	ND	D 1.26	1.14	1	mg/L	124	112	10
		Manganese, Dissolved	EPA 200.7	19110757-010	ND	D 1.07	0.992	1	mg/L	106	99	8
		Nickel, Dissolved	EPA 200.7	19110757-010	ND	D 5.16	4.99	5	mg/L	103	100	3

SPARKS

475 E. Greg Street, Suite 119
 Sparks, Nevada 89431
 tel (775) 355-0202
 fax (775) 355-0817
 EPA LAB ID: NV00925 - ELAP No: 2523

ELKO

1084 Lamoille Hwy
 Elko, Nevada 89801
 tel (775) 777-9933
 fax (775) 777-9933
 EPA LAB ID: NV00926

LAS VEGAS

3230 Polaris Ave. Suite 4
 Las Vegas, Nevada 89102
 tel (702) 475-8899
 fax (702) 622-2868
 EPA LAB ID: NV00932



WETLAB

WESTERN ENVIRONMENTAL TESTING LABORATORY

Specializing in Soil, Hazardous Waste and Water Analysis.

475 E. Greg Street #119 | Sparks, Nevada 89431 | www.WETLaboratory.com
tel (775) 355-0202 | fax (775) 355-0817

1084 Lamoille Highway | Elko, Nevada 89801
tel (775) 777-9933 | fax (775) 777-9933

3230 Polaris Ave., Suite 4 | Las Vegas, Nevada 89102
tel (702) 475-8899 | fax (702) 776-6152

WETLAB Order ID. 19110766

Sparks Control # _____

Elko Control # _____

LV Control # _____

Report

Due Date

Page _____ of _____

Client **Washoe County Community Services Department**

Address **1001 E. 9th Street**

City, State & Zip **Reno, NV 89512**

Contact **Ben Jesch**

Phone **(775) 954-4626**

Collector's Name **Ben Jesch**

Fax **(775) 328-3699**

PWS/Project Name **Swan Lake**

P.O. Number

PWS/Project Number **io 49300**

Email **bjesch@washoecounty.us**

Billing Address (if different than Client Address)

Company _____

Address _____

City, State & Zip _____

Contact _____

Phone _____ Fax _____

Email **bjesch@washoecounty.us**

Turnaround Time Requirements

Standard

5 Day* (25%) 72 Hour* (50%)

48 Hour* (100%) 24 Hour* (200%)

*Surcharges Will Apply

Samples Collected From Which State?

NV CA

Other

Compliance Monitoring?

Yes No

Report Results Via

PDF EDD

Other _____

Report to Regulatory Agency?

Yes No

Standard QC Required?

Yes No

Analyses Requested

SAMPLE ID/LOCATION	DATE	TIME	PRES TYPE *	S	NO. OF CONTAINERS **	Ammonia	BOD / TSS / TDS	Total Nitrogen	Total Phosphorus	E. Coli	Chloride / Sulfate	Nitrate / Total Kjeldahl Nitrogen	Arsenic / Boron	Barium / Iron	Manganese / Nickel	Spl. No.
Compton	0945	11/26/19	Misc.	SW	5	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	
Pompe	11/26/19	1030	Misc.	SW	5	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	
Jean	↓	1010	Misc.	SW	5	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	
Jean DUP	↓	1020	Misc.	SW	5	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	
		11-26 MO														1911 - 6
																0766 4

Instructions/Comments/Special Requirements:

Sample Matrix Key** DW = Drinking Water WW = Wastewater SW = Surface Water MW = Monitoring Well SD = Solid/Sludge SO = Soil HW = Hazardous Waste OTHER: _____

*SAMPLE PRESERVATIVES: 1=Unpreserved 2=H2SO4 3=NaOH 4=HCl 5=HNO3 6=Na2S2O3 7=ZnOAc+NaOH 8=HCl/VOA Vial

Temp	Custody Seal	# of Containers	DATE	TIME	Samples Relinquished By	Samples Received By
1.8 °C	Y N None	20	11/26/19	1115	<i>[Signature]</i>	<i>[Signature]</i>
°C	Y N None					
°C	Y N None					
°C	Y N None					

WETLAB'S Standard Terms and Conditions apply unless written agreements specify otherwise. Payment terms are Net 30.

Client/Collector attests to the validity and authenticity of this (these) sample(s) and, is (are) aware that tampering with or intentionally mislabeling the sample(s) location, date or time of collection may be considered fraud and subject to legal action (NAC445.0636). _____ initial

To the maximum extent permitted by law, the Client agrees to limit the liability of WETLAB for the Client's damages to the total compensation received, unless other agreements are made in writing. This limitation shall apply regardless of the cause of action or legal theory pled or asserted. _____ initial

WETLAB will dispose of samples 90 days from sample receipt. Client may request a longer sample storage time for an additional fee. 301.2E

Please contact your Project Manager for details. _____ initial